

Indira Gandhi Medical College & Research Institute

(Government of Puducherry Institution)

Vazhudavur Road, Kathirkamam, Puducherry – 605 009 Ph :0413-2277545, Fax:0413-2277289

Email: puducherrymedicalcollege@gmail.com, Web: www.igmcricollege.com

No.44-4/IGMC&RI/Pur/Ophthal/e-tender/2018

Date:

Sub: IGMC&RI – Quotation Call letter for IOLs
Surgical Blades – Ophthalmology – Reg.

* * *

The following items are proposed to be procured for this Institution. It is therefore requested to quote the lowest rate for the supply of the same in prescribed format. The quotation shall be submitted by **Hard copy** (As per Circular dated: 31-10-2015) to the Director, IGMC&RI, Puducherry **on or before 24th January 2019**

1. THREE PIECE IOLs - 20 types

2. SURGICAL BLADES – 3 Types

Encl.: List & Specification

Terms and conditions:

1. The supplier should be responsible until the items arrived to this Institution in good condition
2. The rates once quoted and approved will be final and no revision in rates will be allowed
3. Supply should be made within 30 days from the date of receipt of our supply order
4. Delivery should be made on FOR, Puducherry, door delivery basis.
5. The quotation should be based strictly on the specification.
6. **Prices should be inclusive of all taxes.**
7. The material should be supplied strictly in accordance with the specifications mentioned. The items which are not conforming to the specifications should be taken back at their cost.

/ By Order /


ADMINISTRATIVE OFFICER



INDIRA GANDHI MEDICAL COLLEGE AND RESEARCH
INSTITUTE

DEPARTMENT OF OPHTHALMOLOGY

List of 3 Piece IOLs

Sl.No	3 Piece IOL Power	Number Required
1.	16.00D	6
2.	16.50D	6
3.	17.00D	16
4.	17.00D	16
5.	18.00D	40
6.	18.50D	40
7.	19.00D	75
8.	19.50D	75
9.	20.00D	80
10.	20.50D	80
11.	21.00D	100
12.	21.50D	100
13.	22.00D	100
14.	22.50D	100
15.	23.00D	75
16.	23.50D	75
17.	24.00D	25
18.	24.50D	25
19.	25.00D	15
20.	25.50D	10

3 piece Intra Ocular Lens Specifications

- UV absorbing PC lens
- Poly methyl metha acrylate
- 6.0 mm. Biconvex Optic with 2 dialing holes
- Modified 'C' PMMA loop
- 10° angulation
- 13.5mm length
- A constant 118.5

OPHTHALMIC SURGICAL BLADES- Should be sharp and of good quality.

- KERATOME : 2.8mm .angled bevel up – 1000 nos
- CRESCENT : 2.5mm .angled bevel up - 1000 nos
- SIDE PORT : 15° straight -1000 nos

T. Kalyan Prasad
28/12/18